## **Supplemental Technical Specification for**

## **Inspection and Qualification of Asphalt Field Laboratories**

SCDOT Designation: SC-M-404 (07/21)

1.	SCOPE		
1.1	This method covers the process for inspection and qualification of asphalt plant field laboratories for use in testing of asphalt mixtures. This method is not a safety inspection. The Contractor is responsible for maintaining the safety requirements for the asphalt field labs.		
2.	REFERENCED DOCUMENTS		
2.1	SCDOT Standard Specifications	APPROVED: Division Administrator	
2.1.1	Division 300, Division 400		
2.2	AASHTO Standards	Ву:	
2.2.1	R 18	FEDERAL HIGHWAY ADMINISTRATION	
2.3	SCDOT Test Methods		
2.3.1	SC-T-68, SC-T-71, SC-T-75, SC-T-78, SC	C-T-83, SC-T-92, SC-T-96, SC-T-102, SC-T-103	
3.	REQUIREMENTS FOR ALL FIELD L	ABORATORIES	
3.1	Ensure that the building is made and designed for use as either a laboratory or an office that has been converted to serve as a laboratory. The minimum allowable laboratory working floor space is 250 square feet with a minimum inside width of 9 feet. Under no circumstances will a tractor trailer or any other cargo type trailer be acceptable.		
3.2	Provide all required field laboratory equipment listed on the asphalt field lab checklist, SC-M-404A, and ensure that all equipment meets requirements specified in the standard specifications and any supplemental specifications for all mixes.		
3.3	Calibrate the required field lab equipment according to <b>AASHTO R 18</b> , using SCDOT checking, verification, and calibration procedures. Make available all equipment calibration records and keep them in an organized manual in the field laboratory.		
3.4	Equip the laboratory with windows, doors and ventilation systems that function properly and maintain the ambient air temperature between 65°F and 80°F.		

- 3.5 Locate the building in close proximity to the plant and with full view of most major components of the plant. The Asphalt Materials Engineer has the final ruling on what laboratory locations are permissible.
- 3.6 Provide a substantial platform, constructed to the proper height, for use in obtaining asphalt mix samples and inspection of mixtures in truck beds.
- 3.7 Notify the Asphalt Materials Engineer, in writing, when the lab is ready for initial inspection.
- A representative of the Asphalt Materials Engineer will perform an inspection and verify that the lab complies with current standard specifications, and the attached checklist. During the inspection, ensure that a Quality Control Manager or representative is present to certify that all equipment is present and that by signing Form SC-M-404A, they are ensuring that all equipment will remain in the laboratory and will be calibrated or verified as required by **AASHTO R 18**.
- 3.9 Upon meeting all requirements for approval, the lab will be placed or updated monthly on Qualified Product List 76. If at any time all requirements are not met, the labs qualification may be revoked.
- 3.10 After initial inspection, the laboratory will be checked for recertification yearly.

## **I. CONTRACTOR INFORMATION:**

Asphalt Contractor:	Plant Location:	_
Contractor's Representative:		_
Contractor's Signature:		_
Date Inspected:		_
Next Inspection Due Date:	District:	_
II. LAB STRUCTURE YES☑ or NO 図		
Size and type of structure: Floor space	(min. 250 sq ft.)	
The contractor's field laboratory is required to lift the plant cannot be viewed entirely from the amanner where the working components then a monitoring system must be installed separate computer monitor/s that is not to current operating conditions of the plant. The feed and plant production as well as the load	he laboratory, and the view is obstructed in of the asphalt plant are not in plain sight, . The plant monitoring system shall include ed to the plant lab computer to view the e monitor/s will continually monitor the cold	
Is sufficient water available for all tests?		
Is sufficient and satisfactory furniture for off	ice work provided?	
Are satisfactory electric lighting and electric	outlets provided?	
Are suitable worktables and/or benches pro	ovided?	
Are locks provided for the windows and doo	ors?	
Is there sufficient ventilation from solvents a	and other chemicals if applicable?	
Is there a telephone for business use by the	e plant technician or SCDOT personnel?	
Is there Internet access in the laboratory reports?	for e-mail access for file transfer on daily	

## III. EQUIPMENT

1.	Ignition Oven (Meeting requirements of SC-T-75)
	Brand Serial # Model #
2.	Gyratory Compactor (meeting requirements of SC-T-103) including calibration kit (pressure / angle / height / rotation)  a) Four (4) Compaction Molds b) 150mm Compaction Breaking Head for Stability c) 150mm ITS Breaking Head d) 150 mm Gyratory Specimen Protection Paper Discs e) Garden spade minimum 2" wide f) Flat spade 3/4" wide and 6" long g) Extractor jack assembly
	Make: Model/ Serial No:
	Internal angle: Date Last Calibrated:
3.	Compression and Testing Machine – must have a chart recorder to graph stability and flow – minimum capacity of 10,000 lbs. – capable of testing 4" or 6" specimens as specified in SC-T-96 and SC-M-406.  Brand Serial # Model #
4.	Water bath capable of maintaining a constant temperature of $140^{\circ}F \pm 1.8^{\circ}F$ throughout the entire volume of the bath. Water bath meets testing standards specified in SC-T-96. Brand Serial # Model #
5.	Water bath equipped with a water circulator capable of maintaining a constant temperature of 77°F $\pm$ 1.8°F throughout the entire volume of the bath. Water bath meets the testing standards specified in SC-T-68. Brand Serial # Model #
6.	Maximum Gravity Equipment (Ref. SC- T-83):  a) Vacuum pump capable of pulling at least 30mm Hg from daily absolute pressure within 2 minutes of beginning the test.  Brand Serial # Model #  b) Pycnometer or metal container having a capacity of at least 2,000 ml.  c) Ensure that the container has a cover fitted with a rubber gasket and a hose connection. Ensure that the hose opening is covered with a small piece of No. 200 wire mesh to minimize the possibility of loss of fine material.  d) One or more one liter flask to be used as a water vapor trap.  e) Automatic MSG Test Controller to monitor time, vacuum, and vibration operations throughout the test.  f) Kraft paper, or equivalent, for preparation and cooling of sample approximately 3' x 3'
	a) Vibrating Table for constant agitation throughout entire test

7.	Masonry saw equipped with a diamond tip blade and water-cooling system. Ensure that the Masonry saw is capable of slicing a 6 inch diameter core in one pass without disturbing the structure of the core.			
	Brand	_ Serial #	Model #	
8.	cubic feet. Oven is capa <b>Oven</b> .	able of maintaining a	with an inside volume of at least 2.5 temperature of 230°F ± 9°F - <b>Drying</b> Model #	
9.	minimum inside volume temperatures of 265°-3	e of <b>5.0 cubic feet</b> . 25°F). – <b>Mold / Re</b> ł	Oven is capable being set between neating Oven.  Model #	
10.	Two (2) Buckets of ademix from the truck.	quate size (approxin	nately 5 gallons) for sampling asphalt	
11.	Sample quartering table of minimum size 3' x 3' and accessible from at least two sides.			
12.	One (1) Large masonry	r trowel.		
13.	Sample splitter with a n 3 splitter pans.	ninimum of 8 chutes	each 2 inches wide with minimum of	
14.	12" sieve shaker for running HMA gradation samples. (Ro-Tap design or Mary-Ann style) - Must have a tapping device. Also must have the following sieves for the 12 inch shaker: 1", 3/4", 1/2", 3/8", #4, #8, #30, #100, # 200 and bottom pan			
15.	Suitable Sieve Brushes	3		
16	Four (4) certified 50 po	unds weights - <b>SC-</b> 1	-78 only	
17.	Brand	_ Serial #	o 0.1 grams. Model # Model #	
18.	Water softener: No		tener with oil beads	
19.	Cloth Towel – Water absorbent for bulk gravity specimens			
20.	Two (2) calibrated time Brand Brand	_ Serial #	Model # Model #	

21.	Thermometers a) Five (5) Dial Thermometers (50 – 400 degrees Fahrenheit) for plant and				
	road inspectors.	rcury or similar	Thermometer (Such as A	STM 20F or	
	ASTM 45F – NIST	traceable)	·		
	c) One (1) 77°F Mero ASTM 47F – NIST		hermometer (Such as A	ASTM 17F or	
	d) One (1) 300°F	,	(Mercury or Thermoc	ouple- NIST	
	traceable) e) Weather Thermom	neter (measuring	g ambient temperature)		
22.	Penetrating oil or lubrication	on grease for gy	ratory and other equipme	ent	
22					
23.	<ol> <li>A certified caliper readable to 0.01 mm along with an eye comparator with 0.1mm scale (required for verification procedure)</li> </ol>				Ц
24.	. A brass thermometer well (for verification procedure)				
25.	<ol> <li>Cloth sample bags – enough for verification and referee samples (as needed during production)</li> </ol>				
V. CA	ALIBRATION RECORDS (re	ecommend usi	ng AASHTO R18 sched	ule)	
YES 🛭	☑ OR NO ☑				
1.	Ignition oven calibrations fo	r individual job r	mixes posted or filed in th	e field laborator	y? □
2.	2. Ignition oven calibration performed on a monthly basis?				
3.	8. Calibration records available in the field lab?				
4.	4. Equipment calibrations up to date?				
	Equipment	Comment	Equipment	Comment	
	Gyratory Compactor		Vacuum System		
	Gyratory Molds		Thermometers		
	Heating Ovens		Sieves		
	Water Baths		Sieve Shaker		
	Timers		Laboratory Scales		
REMA	ARKS:				
-orm	SC-M-404A		·		